

Product Specification for SS Patient Stretcher Trolley		
Specification विनिर्देश	Sub-Spec उप-विनिर्देश	Value मूल्य
Standards	Conformity to standards	CE
	Certification number and date	1378 & 23/04/2019
	Conformity to Manufacturers quality Standards	ISO 9001 & ISO 13485
Performance Parameters	Overall Length of trolley in mm with 5% tolerance	2030
	Overall width of trolley in mm with 5% tolerance	710
	Material of Stretcher Trolley	SS
	Provision of mattress	Yes
	Thickness of mattress in inch	2
	Length of stretcher in mm with $\pm 5\%$ tolerance	1830
	Width of stretcher in mm with $\pm 5\%$ tolerance	555
	Provision of height adjustment	No
	Provision of Saline rod	Yes
	Material of saline rod	Stainless steel
	Finish of saline rod	Natural
	Provision of storage tray	Yes
	Number of castor wheels	4
	Diameter of castor wheels in mm	125
	Number of Lockable castor wheels	NA
	Material of Castor housing and wheels	High grade non floor-staining synthetic materials with integrated thread guards
	Wheel centre having precision ball bearing to run smoothly	Yes
	Provision of Oxygen cylinder arrangement	Yes
	Provision of railings	No
	Material of railing	Not provided
	Type of railing	Not provided
	Coating on all MS parts	Pre-treated & powder coated
	Polish on all SS parts	finished with Matt polish
Miscellaneous Parameters	All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free	Yes
	Load carrying capacity in kgs	135
	Provision of safety belts	No
	Material of handle	SS
	Urine Bag Holder	No
	Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes